

# Verification of Compliance

No. EC.1282.0B140925.NLMTW42



Certificate's Holder: Ningbo Life Medical Technology Co., Ltd.  
No. 1, Jinghui Road, Hengjie Town, Yinzhou, Ningbo,  
315181, Zhejiang, China

Product: Chamber of Steam Sterilizers  
Model(s): LF08L, LF12L, LF18L, LF22L, LF23L

Technical Data:

Max. Allowable Pressure Ps (bar):	2.39
Nominal size Volume V (L):	8, 12, 18, 22, 23
Range of allowable temperature Ts (°C):	140
Wall thickness (mm):	2

Verification to

Standard:  
EN 764-1:2004, EN 764-5:2002, EN 764-7:2002/AC:2006

related to:  
Directive 97/23/EC (Pressure Equipment)

**Remark:** This Verification of Compliance has been issued on a voluntary basis. ECM confirms that a Technical Construction File (TCF) is existent for the above listed product(s).  
The TCF satisfactorily covers the essential requirements of the above listed standards related to

**Directive 97/23/EC (Pressure Equipment).**

Other relevant Directives have to be observed in case they are applicable.

This Document is only valid for the equipment and configuration described and in conjunction with the TCF detailed above. Whereas the Manufacturer is responsible of the certification of the product(s) and not exempted to perform all the necessary activities before placing the product(s) on the market.

The Manufacturer is also responsible of the internal production control to ensure the product(s) are in compliance with the essential requirements of the above mentioned Directive(s).

This certificate can be checked for validity at [www.entecerma.org](http://www.entecerma.org)

Date of issue 25 SEPTEMBER 2014

Expiry date 24 SEPTEMBER 2019

Certification Chief Manager

Tim Mahan



Certification Deputy Manager

Viola Miller



Ente Certificazione Macchine Srl

Via Ca' Bella, 243 – Loc. Castello di Serravalle – 40053 Valsamoggia (BO) - ITALY

☎ +39 051 6705141 ☎ +39 051 6705156 ☐ info@entecerma.it ☐ www.entecerma.it